



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

***Sharp Precision Calibration, Inc.***  
8570 Dale Street, Centerline, MI 48015

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

**ISO/IEC 17025:2017**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Dimensional Calibration***  
*(As detailed in the supplement)*

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen  
President

<i>Initial Accreditation Date:</i>	<i>Issue Date:</i>	<i>Expiration Date:</i>
January 26, 2005	March 28, 2019	July 31, 2021
<i>Revision Date:</i>	<i>Accreditation No.:</i>	<i>Certificate No.:</i>
July 2, 2020	59322	L19-170-R1

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjllabs.com](http://www.pjllabs.com)*



# Certificate of Accreditation: Supplement

## Sharp Precision Calibration, Inc.

8570 Dale Street, Centerline, MI 48015

Contact Name: Billy Verhey Phone: 586-427-4900

Accreditation is granted to the facility to perform the following calibrations:

### Optical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Optical Comparator Axis Linearity <sup>F</sup>	12 in Maximum	230 $\mu$ in	Glass Master End Measuring Rods Rods Gage Blocks
Optical Comparator Magnification <sup>F</sup>	10 X to 100 X	0.028 % of Magnification	Glass Master Glass Scale
Optical Comparator Axis Squareness <sup>F</sup>	Maximum Travel of Y Axis up to 3 in	170 $\mu$ in	Glass Master
Optical Comparator Chart Angularity <sup>F</sup>	0° to 360°	0.018°	
Toolmakers Microscope Linear Axis <sup>F</sup>	Up to 12 in	200 $\mu$ in	
Toolmakers Microscope Squareness <sup>F</sup>	Maximum Travel of Y Axis up to 3 in	170 $\mu$ in	

### Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Machine Tool Linear Axis <sup>F</sup>	Up to 12 in	920 $\mu$ in	Standard Reference Bar Test Indicator
Surface Plate Flatness <sup>F</sup>	0.5 ft to 25 ft	(75 + 3L) $\mu$ in	Electronic Levels
Surface Plate Repeat Reading <sup>F</sup>	Up to 0.002 in	120 $\mu$ in	Repeat Reading Gage

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor  $k$  (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer<sup>F</sup> would mean that the laboratory performs this calibration at its fixed location.
4. The term L represents length in inches or millimeters as appropriate to the uncertainty statement.